

Department of Biochemistry
NORTH EASTERN HILL UNIVERSITY
Shillong 793022
Meghalaya

Notice Inviting Tender

Ref. No.: DRS-III/Biochem/2015-20

Date: 3rd May, 2017

Sealed quotations are invited from registered and authorized firms/suppliers with reputed track record for the supply and installation of following equipments of established brand in the Department of Biochemistry, NEHU, Shillong, Meghalaya.

Required technical specifications are mentioned below:

Sl. No.	Name of the equipment	Technical Specifications
1	Water purification system	Ultrapure quality water; resistivity 18.2 mΩ; low level of TOC, bacterial count, endotoxin/pyrogen, RNase and DNase and conductivity.
2	Atomic absorption spectrometer	Double beam, Wavelength: 180-900 nm, lamps minimum -4 nos, hollow cathode lamps, auto sampler, lamp turret; automated lamp selection, detection limit in ppb, air compressor; silent, computerized gas control, graphite furnace option, adaptor, burner head:10/5 cm, spectral band width 0.2-2 nm, nebulizer, detector-PMT; hydride system
3	Rotary evaporator	Pump capacity of 10 bar, motorized/automatic lift, chillers, water/oil heating bath; programmable features
4	Flash chromatography system	Capacity/pump: 200 ml/min, Solvent control; Pressure 200 Psi; wavelength: 190-900 nm; fraction collector for different volumes; flow rate: 0-80 ml/min; tube racks bottles: 15-500 ml and accessories (cartridges, columns inclusive).

General Terms and conditions

1. All interested suppliers/ vendors/manufacturers shall send bids to Prof. R. Sharma, Coordinator, UGC-DRS programme, Department of Biochemistry, North-Eastern Hill University, Shillong 793022.

2. The quotation should be submitted on or before **June 5, 2017** through post. Bids received after due date by post will not be considered (the university shall not be responsible for any postal delay).
3. Separate technical and financial bids should be submitted in a sealed envelope.
4. The Departmental Purchase Committee (DPC) reserves the right to accept or reject the quotations received without assigning any reasons to the bidders. No queries in this regard will be entertained.
5. Longer warranty period after installation shall be desirable.
6. The rates should be quoted on FOR basis.
7. The rates should be inclusive of all taxes, freight charges as per norms and should be clearly indicated.
8. Quoted rates should be valid at least for a period of **6 months**.
9. The rates should be quoted along with supporting documents of specifications and technical features.
10. Details of availability of after sales support and warranty period shall have to be furnished during application through post only to address mentioned below.
11. The item should be delivered, installed and free training should be given without additional costs.

Date: 3rd May, 2017

Prof. R. Sharma
Coordinator, UGC SAP-DRS-III
Department of Biochemistry,
NEHU, Shillong